

Member State  
P.R.China

OIML Certificate No.  
R76/1992- CN1-07.01

# OIML CERTIFICATE OF CONFORMITY

## Issuing authority

**Name:** Office of OIML Affairs, General Administration of Quality Supervision, Inspection and Quarantine

**Address:** N0.9 Madiandonglu, Haidian District, Beijing, 100088 China

## Applicant

**Name:** Xiamen Pinnacle Electrical Co., Ltd.

**Address:** 4F, Guangxia Building, Torch High-Tech Zone, Xiamen, Fujian Province, 361006 P.R.China


## Manufacturer of the certified type (if the manufacturer is not the applicant)

Name:

Address:

## Identification of the certified type

### Barcode Label Scale


Type	LS2615E LS2615EC	LS21530E LS21530EC	LS215E LS215EC
Accuracy class			
Max (Max <sub>1</sub> /Max <sub>2</sub> )	6kg/15kg	15kg/30kg	15kg
Min	40g	100g	100g
e (e <sub>1</sub> /e <sub>2</sub> )	2g/5g	5g/10g	5g
n (n <sub>1</sub> /n <sub>2</sub> )	3000/3000	3000/3000	3000
Tare balancing range(subtractive)	≤ 9.995kg	≤ 9.990kg	≤ 7.990kg
Temperature range	0 ~ 40		
Humidity range	10%RH ~ 85%RH		

The numbers of the associated test reports and the included pages:  
NIM-LSmc-0601 that includes 61 pages.

This certificate attests the conformity of the above-mentioned type(represented by the samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML).

R: 76

Edition: 1992

For accuracy class (if applicable): 

This certificate relates only to the metrological and technical characteristics of the type of the instrument concerned ,as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval. The conformity was established by tests described in the associated test reports.

The issuing authority  
  
Office of OIML Affairs  
中国秘书处  
July 25, 2007

The OIML member

  
July 25, 2007

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.