

ČESKÝ METROLOGICKÝ INSTITUT

Member state
Czech Republic



OIML Certificate No.
R76/1992-CZ-06.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Czech Metrology Institute
Address: Okružní 31,
638 00 Brno, CZ
Person responsible: Jan Kalandra

Applicant

Name: Fabricantes de Básculas Torrey, S.A. de C.V.
Address: Ave Los Andes #605 Col. Coyoacán
Monterrey, Nuevo León México 64510
Mexico

Manufacturer of the certified type

Name: Fabricantes de Básculas Torrey, S.A. de C.V.
Address: Ave Los Andes #605 Col. Coyoacán
Monterrey, Nuevo León México 64510
Mexico

Identification of the certified type

Non-automatic weighing instrument, accuracy class III
Type: L-PC series, PCS series

Further characteristics see page 2

This certificate attests the conformity of above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R 76-1, edition 1992, including Amendment 1 (1994)

Member state
Czech Republic

OIML Certificate No.
R76/1992-CZ-06.01


This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation(s) identified above.

This certificate does not bestow any form of legal international approval.


The conformity was established by the results of tests and examinations provided in the associated Test reports: No. 008-05 that includes 34 pages and No. 1013-PT-9023-05 that includes 26 pages and 2 annexes.


Characteristics:

Model	L-PC 20/PCS-20	L-PC 35/PCS-35	L-PC 20T/PCS-20T	L-PC 35/PCS-35T
Max	20 kg	35 kg	20 kg	35 kg
e	0,005 kg	0,010 kg	0,005 kg	0,010 kg
n	4000	3500	4000	3500
Temperature range	-10°C / +40 °C	-10°C / +40 °C	-10°C / +40 °C	-10°C / +40 °C
Tare	9,995 kg	= Max	9,995 kg	= Max


The Issuing Authority
Jan Kalandra

9 March 2006




The OIML Member
Pavel Klenovský

9 March 2006

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and the associated test report is not permitted although either may be reproduced in full.