

# ČESKÝ METROLOGICKÝ INSTITUT

Member state  
Czech Republic



OIML Certificate No.  
R76/1992-CZ-06.02

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Czech Metrology Institute  
Address: Okružní 31,  
638 00 Brno, CZ  
Person responsible: Jan Kalandra

### Applicant

Name: Fabricantes de Básculas Torrey, S.A. de C.V.  
Address: Ave Los Andes #605 Col. Coyoacán  
Monterrey, Nuevo León México 64510  
Mexico

### Manufacturer of the certified type

Name: Dibal, S.A.  
Address: Astintze, 24  
48160 Derio  
Spain

### Identification of the certified type

**Non-automatic weighing instrument, accuracy class III**  
Type: SC-100

Further characteristics see page 2

This certificate attests the conformity of above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

**R 76-1, edition 1992, including Amendment 1 (1994)**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation(s) identified above.


This certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test reports: No. 008-05 that includes 34 pages and No. 1013-PT-9023-05 that includes 26 pages and 2 annexes.


**Characteristics:**

Model	SC-100	SC-100
Max	20 kg	35 kg
e	0,005 kg	0,010 kg
n	4000	3500
Temperature range	-10°C / +40 °C	-10°C / +40 °C
Tare	9,995 kg	= Max

\*Models also available with a customer tower type display

  
**The Issuing Authority**  
Jan Kalandra

9 March 2006



  
**The OIML Member**  
Pavel Klenovský

9 March 2006

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and the associated test report is not permitted although either may be reproduced in full.