

# Český metrologický institut



Member state Czech Republic OIML Certificate No. R76/1992-CZ-09.01

# OIML CERTIFICATE OF CONFORMITY

## **Issuing Authority**

Name:

Czech Metrology Institute

Address:

Okružní 31,

638 00 Brno, CZ

Person responsible: Jan Kalandra

### **Applicant**

Name:

**CAS** Corporation

Address:

#19 Ganap-Ri, Gwangjuk-Myoun,

Yangju-Si, Gyeonggi-Do,

Rep. Of Korea

Manufacturer of the certified type

Name:

**CAS** Corporation

Address:

#19 Ganap-Ri, Gwangjuk-Myoun,

Yangju-Si, Gyeonggi-Do,

Rep. Of Korea

Identification of the certified type

Non-automatic weighing instrument, accuracy class (II) Type: EC series

Further characteristics see page 2

This certificate attests the conformity of above identified type (represented by the sample or samples identified in the associated test report) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R 76-1, edition 1992, including Amendment 1 (1994)

Page 1 This certificates includes 2 pages

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation(s) identified above.

This certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test reports: No. 0115-UL-N041-08 that includes 31 pages, No. 1013-PT-9006-08 that includes 24 pages including 2 annexes and No. 8551-PT-E095-09 that includes 2 pages and 2 annexes.

#### Characteristics:

| Max                           | 3000g                                     | 6000g | 15000g | 30000g |
|-------------------------------|---|-------|--------|--------|
| Min                           | 20g                                       | 40g   | 100g   | 200g   |
| Verification scale interval e | 1g  | 2g    | 5g     | 10g    |
| Tare (subtractive)            | - Max                                     |       |        |        |
| Temperature range             | 0°C to +40°C                              |       |        |        |
| Power supply                  | AC Adaptor DC 12V/1A or 12V/800mA         |       |        |        |
|                               | Internal rechargeable sealed acid battery |       |        |        |
| Accuracy class                |   |       |        |        |

The Issuing Authority Jan Kalandra

30 June 2009

Session of the sessio

The CIML Member Pavel Klenovský

30 June 2009

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate is issued, partial quotation of the certificate and the associated test report is not permitted although either may be reproduced in full.