

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R76/1992-DE1-05.06

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Roman Schwartz

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen
GERMANY

Manufacturer of the certified type is the applicant.

Identification of the certified type

Nonautomatic electromechanical weighing instrument
Types: SC(-H) 100, SC(-H) 200, SC(-H) 400,
SC(-H) 500, SC(-H) 800, SC-C

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 1992, including Amendment 1 (1994),
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R76/1992-DE1-05.06

The conformity was established by the results of tests and examinations provided in the associated report No. 1.12-4019152 (9 pages) and Test Reports No. 1.12-4019152/1 (53 pages) and No. 1.12-4019152/2 (86 pages).

The Issuing Authority

Dr. R. Schwartz
Direktor und Professor

14.10.2005

The OIML Member

Prof. Dr. M. Kochsiek
Vizepräsident

14.10.2005

Identification of the pattern (continued)

The weighing instrument consists of a weighing platform with one strain-gauge load cell and of an incorporated indicating device for displaying the weighing results, and of a keypad to operate the instrument.

The weighing ranges with Max, Min, e, d and number of verification scale intervals may be chosen within the limits of No. 3.2 of R 76-1 and of the table 1.

Table 1

Accuracy class	III
Type	SC...
Max	≤ 300 kg
e =	1 g ... 50 g
d =	1 g ... 50 g
n ≤	6000
Tare balancing range (subtr.)	≤ 100 % of Max
Preset tare range	≤ Max ≤ Max ₁ at multi-interval instruments
Temperature range	-10 °C / 40 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.