

Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML
Germany



OIML Certificate N°
R76/1992-DE1-07.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: Physikalisch-Technische Bundesanstalt
Address: Bundesallee 100, 38116 Braunschweig
Person responsible: Dr. Panagiotis Zervos

Applicant

Name: Bizerba GmbH & Co. KG
Address: Wilhelm-Kraut-Str. 65, 72336 Balingen

Manufacturer of the certified type is the applicant.

Identification of the certified type

Nonautomatic electromechanical weighing instrument
Type: ECII...

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, edition 1992, including Amendment 1 (1994),
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

Physikalisch-Technische Bundesanstalt

OIML Certificate N°
R76/1992-DE1-07.02

The conformity was established by tests described in the associated Report 1.12-4028474 (6 pages) and the Test Report No. 1.12-4028474/1 (53 pages).

The Issuing Authority

Dr. P. Zervos
Regierungsdirektor

15.03.2007

The OIML Member

Dr. R. Schwartz
Direktor und Professor

15.03.2007

Identification of the pattern (continued)

The weighing instrument consists of a weighing platform with one strain-gauge load cell of an incorporated indicating device for displaying the weighing results, and of a keypad to operate the instrument.

The weighing ranges with Max, Min, e, d and number of verification scale intervals may be chosen within the limits of No. 3.2 of R 76-1 and of Table 1.

Table 1

Accuracy class	III
Type	ECII...
Max	≤ 30 kg
e	= 1g ... 10g
d	= 1g ... 10g
n	≤ 6000
Tare balancing range (subtr.)	≤ 100 % of Max
Preset tare range	≤ Max ≤ Max ₁ at multi-interval instruments
Temperature range	-10 °C / 40 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.