

# Physikalisch-Technische Bundesanstalt

Braunschweig und Berlin

Member State of OIML  
Germany



OIML Certificate N°  
**R76/1992-DE1-07.09**

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: Physikalisch-Technische Bundesanstalt  
Address: Bundesallee 100, 38116 Braunschweig  
Person responsible: Dr. Panagiotis Zervos

### Applicant

Name: ERTE Endustriyel Elektronik Sanayi ve Ticaret Limited Sirketi  
Address: Sakarya Cad. 142/A, 35330 BALCOVA  
TURKEY

Manufacturer of the certified type is the applicant.

### Identification of the certified type

Non-automatic electromechanical weighing instrument with or without lever system

Type: E200

Further characteristics see page 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, edition 1992, including Amendment 1 (1994),  
for accuracy class(es) (III) (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

# Physikalisch-Technische Bundesanstalt

OIML Certificate N°  
**R76/1992-DE1-07.09**

The conformity was established by the results of tests and examinations provided in the associated Report No. 1.12-4030759 (14 pages) and the Test Report No. 1.12-4030759/1 (44 pages).

## The Issuing Authority

Dr. P. Zervos  
Direktor und Professor

13.11.2007

## The CIML Member

Dr. R. Schwartz  
Direktor und Professor

13.11.2007

Identification of the type (continued):

Designed as platform-, hopper-, wall-mounted-, crane- or overhead track scale, with or without lever system. The weighing instrument is equipped with a lever system or directly introduces the force into one or more load cells.

The weighing ranges comprising Max, verification scale intervals, number of verification scale intervals and scale intervals may be selected considering the limiting values in table 1.

Table 1:

Accuracy class	(III)	(IIII)
Max	0.3 kg to 400 t	
Number of verification intervals $n \leq$ <sup>1)</sup>	6000	1000
Temperature range	-10 °C to +40 °C	
Tara-balancing range	$\leq$ 100 % of Max (subtractive)	
Preset tare range	$\leq$ 100 % of Max <sup>1)</sup>	

<sup>1)</sup> this applies to each range of single- and multiple range instruments

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report(s) is not permitted, although either may be reproduced in full.