



Member State  
Denmark

OIML Certificate N°  
R76/1992-DK3-11.01

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgaard Jensen

### Applicant

Name: **Tüm Elektronik Mühendislik San. ve Tic. Ltd.Şti.**  
Address: İstanbul Deri Organize San. Bölgesi, 1. Yol, H7 Parsel,  
Orhanlı  
34956 Tuzla, İstanbul  
TURKEY

### Manufacturer

of the certified pattern: **Tüm Elektronik Mühendislik San. ve Tic. Ltd.Şti.**

### Identification

of the certified pattern: **Non-automatic weighing instrument**  
**Type: TEB / MEB**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 1992, including amendment 1 (1994)**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages

DELTA



## OIML Certificate N° R76/1992-DK3-11.01

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-199788, dated 2005-12-21, including 101 pages.

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
15 February 2011

J. Hovgaard Jensen  
Certification Officer

### Characteristics

Accuracy class		III
Weighing range		Single-interval or 2 multi-ranges
Maximum capacity	Max	$\leq 375000$ kg
Minimum capacity	Min	20 e
Verification scale interval	e	$\geq 1$ g
Number of intervals	n	$\leq 3000$ (per weighing range)
Temperature range for the indicator		-10 °C / +40 °C

### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.