



Member State  
Denmark

OIML Certificate N°  
R76/1992-DK3-11.02

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: **DELTA**  
Address: Venlighedsvej 4  
2970 Hørsholm  
Denmark

Person responsible: J. Hovgaard Jensen

### Applicant

Name: **Tunaylar Baskül Sanayi ve Ticaret A.Ş.**  
Address: Akçaburgaz Mah. 88 Sok. No:7  
Esenyurt - İstanbul  
Turkey

### Manufacturer

of the certified pattern: **Tunaylar Baskül Sanayi ve Ticaret A.Ş.**

### Identification

of the certified pattern: **Non-automatic weighing instrument**  
**Type: Load Line-3**  
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
**edition 1992, including amendment 1 (1994)**  
**for accuracy class III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



## OIML Certificate N° R76/1992-DK3-11.02

The conformity was established by tests described in the associated test reports from DELTA, DK, no. DANAK-199737 that includes 10 pages and 1 annex (64 pages).

The issuing authority: **DELTA, OIML Issuing Authority DK3**  
21 February 2011

J. Hovgaard Jensen  
Certification Officer

### Characteristics

Accuracy class		III
Weighing range		Single-interval, multi-range, or multi-interval
Number of intervals	Max / e	≤ 7,500 (class III)
Minimum input voltage	$\Delta U_{\min}$	0.33 $\mu$ V per verification scale interval
Measuring range	Min. / Max.	-10 mV / 40 mV
Excitation voltage	$U_{\text{exc}}$	5 VDC
Load cell impedance	Min. / Max.	42 ohm / 1,200 ohm
Load cell connecting system		6 wire system, screened
Module fractional factor	$p_i$	0.5 for the indicator
Interface		Protective, according to paragraph 5.3.6
Connected load cells		Shall comply with R60
Temperature range for the indicator		-10 °C / +40 °C

#### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.