



Member State
Denmark

OIML Certificate N°
R76/1992-DK3-11.07

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **DELTA**
Address: Venlighedsvej 4
2970 Hørsholm
Denmark

Person responsible: J. Hovgård Jensen

Applicant

Name: **TScale Electronics Mfg. (Kunshan) Co., Ltd.**
Address: No. 99 Shunchang Road,
Zhoushi, Kunshan, Jiangsu
CHINA

Manufacturer

of the certified pattern: **TScale Electronics Mfg. (Kunshan) Co., Ltd.**

Identification

of the certified pattern: **Non-automatic weighing instrument**
Type: S29 / SW
Further characteristics are set out on page 2.

This certificate attests the conformity of the above mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
edition 1992
for accuracy class III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

Page 1. This certificate includes 2 pages



OIML Certificate N° R76/1992-DK3-11.07

The conformity was established by tests described in the associated test reports from DELTA, DK, No. DANAK-1910384, dated 13-01-2009 that includes 26 pages, and DELTA, DK, No. DANAK-1910465, dated 01-04-2009 that includes 45 pages DELTA, DK, No. DANAK-1910594, dated 14-08-2009 that includes 36 pages DELTA, DK, No. DANAK-1911100, dated 15-10-2010 that includes 42 pages

The issuing authority: **DELTA, OIML Issuing Authority DK3**
16 November 2011

J. Hovgård Jensen
Certification Officer

Characteristics

Accuracy class		III	
		Single-interval	multi-interval
Weighing range			
Maximum capacity	Max	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$	
Number of intervals	n_i	≤ 3000	$\leq 2 \times 3000$
Minimum capacity	Min	$20e$	$20e_1$
Verification scale interval	e	$e \geq 1 \text{ g}$	$e_i \geq 1 \text{ g}$
Maximum tare effect	T	$\leq -\text{Max}$	
Temperature range		$-10 \text{ °C} / +40 \text{ °C}$	

Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.