



Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland

OIML Certificate No  
R76/1992-GB1-04.07  
Revision 1

## OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**  
Address: **Stanton Avenue  
Teddington  
Middlesex  
TW11 0JZ  
United Kingdom**

Person responsible: **Paul Dixon – Product Certification Manager**

Applicant

Name: **Avery Berkel**  
Address: **Foundry Lane  
Smethwick  
West Midlands B66 2LP  
United Kingdom**

Identification of the certified pattern:

**IX series**  
**Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

<b>OIML:</b>	<b>R76</b>
<b>Edition:</b>	<b>1992 (E)</b>
<b>Accuracy class:</b>	<b>III</b>

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated Test Reports:

Test report: AWTX 00157 having 46 pages  
Test report: NWML 00904 having 9 pages  
Pattern evaluation checklist: F20076 which includes 12 pages

The issuing authority



Mr P R Dixon

The CIML member



Mr P E Mason

Date 04 December 2008

Ref: T1128/0180

This revision replaces earlier versions of this certificate.

### Characteristics:

Mains powered, Class III, non-automatic weighing instrument designated the IX Series.

Maximum capacity	6 kg ≤ Max ≤ 15 kg
Number of scale intervals	≤ 6000 for Class III with single interval
	≤ 3000 for Class III with multi-interval, maximum two partial weighing ranges
Verification scale interval	e ≥ 1 g
Operating temperature range	-10 °C to + 40 °C
Power supply	220 - 240 V ac 50/60 Hz
Load cell	Avery T702

### CERTIFICATE HISTORY

ISSUE NO.	DATE	DESCRIPTION
R76/1992-GB1-04.07	22 December 2004	Type approval first issued.
R76/1992-GB1-04.07 Revision 1	04 December 2008	Company name changed from Avery Weigh-Tronix Ltd to Avery Berkel.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.