



**Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland**

**OIML Certificate No  
R76/1992-GB1-07.06  
Revision 1**

## **OIML CERTIFICATE OF CONFORMITY**

Issuing authority

Name: **National Weights and Measures Laboratory**  
Address: **Stanton Avenue  
Teddington  
Middlesex  
TW11 0JZ  
United Kingdom**

Person responsible: **Paul Dixon  
Product Certification Manager**

Applicant

Name: **CAS Corporation**  
Address: **19 Ganap-Ri  
Gwangjuk-Myoun  
Yangji-Si  
Gyeonggi-Do 482-841  
Republic of Korea**

Manufacturer of the certified pattern is the Applicant, or the following companies:

Shanghai CAS Electronics Co., Ltd,  
Maixinroad 448, Xinqiaozhen, Songjiangqu,  
Shanghai, China  
CAS Elektronik San. Tic. A.S.  
Yukari Dudulu, Bostanci Cad. Mevdudi Sokak No: 34  
Umraniye-Istanbul / Turkey

Identification of the certified pattern:

**Dolphin Series non-automatic weighing instruments  
Further characteristics see page 2**

**OIML Certificate No  
R76/1992-GB1-07.06  
Revision 1**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML:** R76  
**Edition:** 1992 (E)  
**Accuracy class:** III

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports Number SN: 1022 having 44 pages and SN: 1032 having 15 pages, and associated pattern evaluation checklist G20121 having 12 pages.

The issuing authority



Mr P R Dixon

The CIML member



Mr P Mason

Date: 23 October 2007  
Ref: T1128/0115

Characteristics: Class III, mains or battery-operated, self-indicating, dual-interval, non-automatic weighing instruments designated Dolphin Series and comprising the EB, DB-II and DB-1H models and their variants

Max	30/60 kg	60/150 kg
Min	200 g	400 g
e	10/20 g	20/50 g
T	-29.99 kg	-59.98 kg

Main features:

- Steel construction
- Operator's keypad
- Stainless steel load receptor
- Pole-mounted display
- Level indicator

**OIML Certificate No  
R76/1992-GB1-07.06  
Revision 1**

Devices:

- Initial zero setting device ( $\leq 20\%$  of Max)
- Semi-automatic zero setting device ( $\leq 4\%$  of Max)
- Zero tracking device ( $\leq 4\%$  of Max)
- Zero indicator
- Net indicator
- Semi-automatic subtractive tare balancing device
- Gravity compensation
- Counting / check weighing / weighing unstable samples modes (DB-II Model)
- Price-computing (EB Model)

This revision replaces earlier versions of this certificate.

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.