



Member State of OIML
United Kingdom of Great Britain
and Northern Ireland

OIML Certificate No
R76/1992-GB1-07.07
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon
Product Certification Manager**

Applicant

Name: **Huntleigh Healthcare Ltd**
Address: **Unit 3
Trident Drive
Britannia Park
Wednesbury
West Midlands
WS10 7XB
United Kingdom**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**Huntleigh Healthcare Enterprise 9000 or Enterprise 9100
hospital bed with weighing facility
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML: R76
Edition: 1992 (E)
Accuracy class: III**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test reports: SN: 01012 having 43 pages, SN1073 having 32 pages, and associated pattern evaluation checklist P00089 having 12 pages.

This revision replaces previous versions of the certificate.

The issuing authority



Mr P R Dixon

The CIML member



Mr P Mason

Date: 29 October 2008

Ref: T1128/0173

Characteristics: The Huntleigh Healthcare Enterprise 9000 and Enterprise 9100 hospital beds incorporate Class III non-automatic weighing instruments have the following characteristics:

- Max 250 kg
- Min 10 kg
- $e = 0.5$ kg
- Weighing control buttons and weight displays in both safety sides.
- Four Vishay Tedeo Huntleigh model 3410 load cells, each having a capacity of 1000 lb; or four Zemic shear beam load cell model H8C, each having a capacity of 500 kg.
- "AUTO_COMP" function, which allows the weight value to be frozen while medical equipment is added or removed, without disturbing the patient.
- Extended indicating device temporarily giving interval of 0.1 kg.
- Operating temperature range +10°C to +40°C.
- Optional local gravity adjustment
- Alternative construction: : the headboard is mounted off the mattress platform instead of the base-frame

The Enterprise 9000 and Enterprise 9100 have identical weighing equipment, but different medical facilities.

The following alternative construction is also authorised: the headboard is mounted off the mattress platform instead of the base-frame.

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.