



**Member State of OIML  
United Kingdom of Great Britain  
and Northern Ireland**

**OIML Certificate No  
R76/1992-GB1-07.07**

## **OIML CERTIFICATE OF CONFORMITY**

Issuing authority

Name: **National Weights and Measures Laboratory**  
Address: **Stanton Avenue  
Teddington  
Middlesex  
TW11 0JZ  
United Kingdom**

Person responsible: **Paul Dixon  
Product Certification Manager**

Applicant

Name: **Huntleigh Healthcare Ltd**  
Address: **Woden Road West  
Wednesbury  
West Midlands  
WS10 7BL  
United Kingdom**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**Huntleigh Healthcare Enterprise 9000 or Enterprise 9100  
hospital bed with weighing facility  
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test reports) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**OIML: R76  
Edition: 1992 (E)  
Accuracy class: III**

**OIML Certificate No  
R76/1992-GB1-07.07**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report: SN: 01012 having 43 pages, and associated pattern evaluation checklist G20107 having 12 pages.

The issuing authority



Mr P R Dixon

Date: 14 August 2007

Ref: T1128/0078

The CIML member



Dr J W Llewellyn

Characteristics: The Huntleigh Healthcare Enterprise 9000 and Enterprise 9100 hospital beds incorporate Class III non-automatic weighing instruments have the following characteristics:

- Max 250 kg
- Min 10 kg
- $e = 0.5$  kg
- Weighing control buttons and weight displays in both safety sides.
- Four Vishay TedeA Huntleigh model 3410 load cells, each having a capacity of 1000 lb.
- "AUTO\_COMP" function, which allows the weight value to be frozen while medical equipment is added or removed, without disturbing the patient.
- Extended indicating device temporarily giving interval of 0.1 kg.
- Operating temperature range  $+10^{\circ}\text{C}$  to  $+40^{\circ}\text{C}$ .

The Enterprise 9000 and Enterprise 9100 have identical weighing equipment, but different medical facilities.

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.