



**Member State of OIML
United Kingdom of Great Britain
and Northern Ireland**

**OIML Certificate No
R76/1992-GB1-08.01**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon
Product Certification Manager**

Applicant

Name: **Fabricantes de Basculas Torrey, S.A. de C.V.**
Address: **Ave. Los Andes 605
Col. Coyoacan C.P. 64510
Monterrey NL
Mexico**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**Torrey TLS-20 and TLS-40
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

**OIML Certificate No
R76/1992-GB1-08.01**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report SN1015 having 46 pages and the associated pattern evaluation checklist G20105 having 12 pages.

The issuing authority



Mr. P. Dixon

The CIML member



Mr. P. Mason

Date: 25 January 2008

Reference: T1128/0091

Characteristics: Class III non-automatic weighing instrument designated the Torrey touch screen TLS-20 and TLS-40.

Instrument	TLS-20	TLS-40
Maximum capacity	20 kg	40 kg
Minimum load	100 g	200 g
Verification scale interval	5 g	10 g
Operating temperature range	-10 °C to + 40 °C	
Power supply	240 V ac 50/60 Hz	
Load cell	Medidata model M022	

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.