



**Member State of OIML
United Kingdom of Great Britain
and Northern Ireland**

**OIML Certificate No
R76/1992-GB1-08.02**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon
Product Certification Manager**

Applicant

Name: **Fabricantes de Basculas Torrey, S.A. de C.V.**
Address: **Ave. Los Andes 605
Col. Coyoacan C.P. 64510
Monterrey NL
Mexico**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**Torrey L-EQ and EQ
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

**OIML Certificate No
R76/1992-GB1-08.02**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated:

Test reports: SN: 1039 having 41 pages
Pattern evaluation checklist: G20119 having 12 pages

The issuing authority

The CIML member



Mr P R Dixon



Mr P Mason

Date: 27 February 2008

Characteristics: Class III non-automatic weighing instrument designated the Torrey L-EQ and EQ.

The individual models are as follows:

Model	Max	Min	e =	n
L-EQ 4	4 kg	0.020 kg	0.001 kg	4000
L-EQ 8	8 kg	0.040 kg	0.002 kg	4000
L-EQ 10	10 kg	0.100 kg	0.005 kg	2000
L-EQ 20	20 kg	0.100 kg	0.005 kg	4000
EQ-4	4 kg	0.020 kg	0.001 kg	4000
EQ-8	8 kg	0.040 kg	0.002 kg	4000
EQ-10	10 kg	0.100 kg	0.005 kg	2000
EQ-20	20 kg	0.100 kg	0.005 kg	4000

The L-EQ and EQ use the same electronics and load cells. However, their housings, although both stainless steel, are different, and the EQ has fewer facilities.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.