



**Member State of OIML
United Kingdom of Great Britain
and Northern Ireland**

**OIML Certificate No
R76/1992-GB1-09.01**

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: **National Weights and Measures Laboratory**
Address: **Stanton Avenue
Teddington
Middlesex
TW11 0JZ
United Kingdom**

Person responsible: **Paul Dixon
Product Certification Manager**

Applicant

Name: **LESÁK-ZEMAN s.r.o.**
Address: **Vranovská 699/33
61400 Brno
Czech Republic**

Manufacturer of the certified pattern is the Applicant.

Identification of the certified pattern:

**KS1-XX
Further characteristics see page 2**

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML:	R76
Edition:	1992 (E)
Accuracy class:	III

**OIML Certificate No
R76/1992-GB1-09.01**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests and examination described in the associated:

Test reports: SN: 1039 having 41 pages
Pattern evaluation checklist: G20119 having 12 pages

The above reports are associated to the Torrey L-EQ and EQ instruments. The KS1-XX instruments share the same components, but have a slightly different overlay cover.

The issuing authority

The OIML member



Mr P R Dixon
for NWML

Mr P Mason

Date: 25 February 2009

Ref: T1128/0181

Characteristics: Class III non-automatic weighing instrument designated the KS1-XX, where XX relates to the capacity of the instrument.

The individual models are as follows:

Model	Max	Min	e =	n
KS1-4	4 kg	0.020 kg	0.001 kg	4000
KS1-8	8 kg	0.040 kg	0.002 kg	4000

CERTIFICATE HISTORY

ISSUE NO.	DATE	DESCRIPTION
R76/1992-GB1-09.01	25 February 2009	Certificate first issued.

Important note: Apart from the mention of the certificates reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.