


OIML Member State
Japan



OIML Certificate N°
R76/1992-JP1-06.05



OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Hiroyuki Yoshikawa, President of AIST

Applicant

Name: TANITA CORPORATION
Address: 14-2, 1chome, Maeno-cho, Itabashi-ku, Tokyo, 174-8630, Japan

Manufacturer of the certified type

Name: TANITA AKITA CORPORATION
Address: 28-1, Shimotamogizoe, Horiminai, Aza, Daisen-shi, Akita-ken,
014-0113, Japan

Identification of the certified type:

Type MC-180MA
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment1 (1994),
for accuracy class: **III**


This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.

OIML Member State
Japan



OIML Certificate N°
R76/1992-JP1-06.05



The conformity was established by the results of tests and examinations provided in the associated Test Report:
N° 06-05/R76 that includes 52 pages;

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. H. Yoshikawa
President of AIST

A handwritten signature in cursive script.

Dr. Y. Miki

2006-08-29

2006-08-29

Characteristics:

Type	MC-180MA
Class	III
Max = (Max ₁ / Max ₂)	200 kg / 270 kg
e= (e ₁ / e ₂)	0.05 kg / 0.1 kg
n= (n ₁ / n ₂)	4000 / 2700
Min	1 kg
Maximum Tare-balancing effect	T = - 10 kg
Temperature range	+ 5 °C / +35 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.