



OIML Member State  
Japan



OIML Certificate N°  
R76/1992-JP1-07.01

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: National Metrology Institute of Japan /National Institute of Advanced Industrial Science and Technology (NMIJ/AIST)  
Address: AIST Tsukuba Central 3-9  
Tsukuba Ibaraki 305-8563, Japan  
Person responsible: Dr. Hiroyuki Yoshikawa, President of AIST

### Applicant

Name: A&D Company, Limited  
Address: 3-23-14, Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013,  
JAPAN

### Manufacturer of the certified type

Name: Litra Co.,Ltd  
Address: 7-5, Harajyuku, Hidaka-shi, Saitama 350-1205, JAPAN

### Identification of the certified type:

Type AD-6121A  
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, Edition 1992, including amendment 1 (1994),  
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State  
Japan

OIML Certificate N°  
R76/1992-JP1-07.01

The conformity was established by the results of tests and examinations provided in the associated Evaluation Report N° R76/1992-JP1-07.01-ER.

The Issuing Authority  
NMIJ/AIST

The OIML Member

Dr. H. Yoshikawa  
President of AIST  
2007-02-15



Dr. Y. Miki  
2007-02-15

Characteristics:

Type	AD-6121A
Class	III
Max =	150 kg
e =	0.1 kg
n =	1500
Min =	2 kg
Maximum Tare-balancing effect	T = - 150 kg
Temperature range	+ 5 °C / + 35 °C

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.