



OIML Member State  
Japan



OIML Certificate N°  
R76/1992-JP1-09.02

## OIML CERTIFICATE OF CONFORMITY

### Issuing Authority

Name: National Metrology Institute of Japan /National Institute of  
Advanced Industrial Science and Technology (NMIJ/AIST)  
Address: AIST Tsukuba Central 3-9  
Tsukuba Ibaraki 305-8563, Japan  
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

### Applicant

Name: Ishida Co., Ltd.  
Address: 44 Sanno-cho, Shogoin, Sakyo-ku, Kyoto-city, 606-8392, Japan

### Manufacturer of the certified pattern

Name: Ishida Co., Ltd.  
Address: 44 Sanno-cho, Shogoin, Sakyo-ku, Kyoto-city, 606-8392, Japan

### Identification of the certified type:

Non-automatic weighing instruments  
Type: ITX-..., ITB-..., ATX-..., ATB-... series  
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76-1**, Edition 1992, including amendment 1 (1994),  
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State  
Japan

OIML Certificate No  
R76/1992-JP1-09.02

The conformity was established by the results of tests and examinations provided in the associated Evaluation Report N° 21-003.

The Issuing Authority  
NMIJ/AIST

The OIML Member

Dr. T. Nomakuchi  
President of AIST  
2009-10-27



Dr. Y. Miki  
2009-10-27

Characteristics:

|                                |   |
|--------------------------------|---|
| Type                           | ITX-..., ITB-..., ATX-..., ATB-... series |
| Class                          | III                                       |
| Max                            | 1.5 kg ... 150 kg                         |
| e                              | 0.5 g ... 50 g                            |
| n <sup>1)</sup>                | ≤ 3000                                    |
| Min                            | ≥ 10 g                                    |
| Maximum Tare-balancing(subtr.) | ≤ Max                                     |
| Temperature range              | - 5 °C / + 35 °C                          |

<sup>1)</sup> Valid for range of single- and multiple range instruments

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.