



OIML Member State
Japan



OIML Certificate N^o
R76/1992-JP1-10.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: TANITA CORPORATION
Address: 14-2, 1chome, Maeno-cho, Itabashi-ku, Tokyo, 174-8630, Japan

Manufacturer of the certified pattern

Name: TANITA AKITA CORPORATION
Address: 28-1, Shimotamogizoe, Horiminai, Aza, Daisen-shi, Akita-ken,
014-0113, Japan

Identification of the certified type:

Non-automatic weighing instruments
Type: TL-260A, TL-280A
Non-automatic weighing instruments (Instruments for direct sales to the public)
Type: TL-270A, TL-290A
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment 1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan

OIML Certificate No
R76/1992-JP1-10.01

The conformity was established by the results of tests and examinations provided in the associated:

Evaluation Report N° 21-014(test report N° 10-01-A/R76:1992 that includes 59 pages.
test report N° 10-01-B/R76:1992 that includes 35 pages.)

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. T. Nomakuchi
President of AIST

Dr. Y. Miki

2010-03-26

2010-03-26

Characteristics:

Type ¹⁾	TL-260A, TL-280A	TL-270A, TL-290A
Class	III	III
Max	4 kg ... 15 kg	4 kg ... 15 kg
e	1 g ... 5 g	1 g ... 5 g
d	0.2 g ... 1 g	—
n	≤ 4000	≤ 4000
Min	≥ 4 g	≥ 20 g
Maximum Tare-balancing(subtr.)	≤ Max/2	≤ Max/2
Temperature range	0 °C / + 35 °C	0 °C / + 35 °C

¹⁾ TL-270A, TL-290A (Instruments for direct sales to the public)

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.