



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-12.01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: A&D Company, Limited
Address: 3-23-14 Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN

Manufacturer of the certified type

Name: A&D SCALES CO.,LTD
Address: 162-4 INSAN-NI,DEOGSAN-MYEON,JINCHEON-GUN,
CHUNGCHONGBUG-DO,KOREA

Identification of the certified type

Non-automatic weighing instruments
Type: SC/SE series
Further characteristics see page 3

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment 1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan

OIML Certificate No
R76/1992-JP1-12.01

The conformity was established by the results of tests and examinations provided in the associated:

Evaluation Report N° 24-01(Test Report N° 11-02/R76:1992 that includes 58 pages.)

The Issuing Authority

NMIJ/AIST



Dr. T. Nomakuchi

President of AIST

2012-03-09

The OIML Member

Dr. Y. Miki

2012-03-09



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-12.01

Characteristics:

Type	SC-30KAM SC-30KBM SE-30KAM SE-30KBM	SC-60KAM SC-60KBM SC-60KAL SC-60KBL SE-60KAM SE-60KBM SE-60KAL SE-60KBL	SC-150KAM SC-150KBM SC-150KAL SC-150KBL SE-150KAM SE-150KBM SE-150KAL SE-150KBL
Class	III		
Max	30 kg	60 kg	150 kg
e	10 g	20 g	50 g
n	3000		
Min	200 g	400 g	1 kg
Tare-balancing range	≤ 100 % of Max		
Temperature range	-10 °C / +40 °C		
Power supply	Dry battery		

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.