



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-12.02

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Tamotsu Nomakuchi, President of AIST

Applicant

Name: Unipulse Corporation
Address: 9-11 Nihonbashi Hisamatsucho, Chuo-ku, Tokyo 103-0005, Japan

Manufacturer of the certified type

Name: Unipulse Corporation Technical Center
Address: 1-3 Sengendainishi, Koshigaya, Saitama 343-0041, Japan

Identification of the certified type

Non-automatic weighing instruments
Type: F701-S
Further characteristics see page 2

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment 1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan

OIML Certificate N°
R76/1992-JP1-12.02

The conformity was established by the results of tests and examinations provided in the associated:

Evaluation Report N° 24-02 (Test Report N° 12-01/R76:1992 that includes 56 pages.)

The Issuing Authority

The OIML Member

NMIJ/AIST



Dr. T. Nomakuchi

Dr. Y. Miki

President of AIST

2012-08-09

2012-08-09

Characteristics:

Type	F701-S		
Class	III		
Max	120 kg		
e	0.02 kg	0.05 kg	0.1 kg
n	6 000	2 400	1 200
Min	0.40 kg	1.0 kg	2.0 kg
Tare-weighing rang	≤ 100 % of Max		
Temperature range	-10 °C ~ +40 °C		
Power supply	AC 100 V~240 V, 50/60 Hz		

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.