



OIML Member State
Japan

OIML Certificate N^o
R76/1992-JP1-14. 01

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: National Metrology Institute of Japan /National Institute of
Advanced Industrial Science and Technology (NMIJ/AIST)
Address: AIST Tsukuba Central 3-9
Tsukuba Ibaraki 305-8563, Japan
Person responsible: Dr. Ryoji Chubachi, President of AIST

Applicant

Name: A&D Company, Limited
Address: 3-23-14 Higashi-ikebukuro, Toshima-ku, Tokyo 170-0013, JAPAN

Manufacturer of the certified type

Name: A&D ELECTRONICS (Shen Zhen) CO., LTD.
Address: Tantou Datianyang Industrial Eatate, Songgang Town, Baoan
District, Shenzhen, China

Identification of the certified type

Non-automatic weighing instruments
Type: SJ-WP series
Further characteristics see page 3

This certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Evaluation Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76-1, Edition 1992, including amendment 1 (1994),
for accuracy class: **III**

This certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This certificate does not bestow any form of legal international approval.



OIML Member State
Japan


OIML Certificate No
R76/1992-JP1-14. 01

The conformity was established by the results of tests and examinations provided in the associated:

Evaluation Report N° 26-01(Test Report N° 14-01-A / R76:1992 that includes 36 pages and Test Report N° 14-01-B / R76:1992 that includes 37 pages.)

The Issuing Authority

NMIJ/AIST



Dr. R. Chubachi

President of AIST

2014-10-23

The CIML Member

Dr. Y. Miki

2014-10-23



OIML Member State
Japan

OIML Certificate N^o
R76/1992-JP1-14.01

Characteristics:

Type	SJ-3000WP	SJ-6000WP	SJ-15KWP	SJ-30KWP
Class	(III)			
Max	1 500 / 3 000 g	3 000 / 6 000 g	6 / 15 kg	15 / 30 kg
e	0.5 / 1 g	1 / 2 g	0.002 / 0.005 kg	0.005 / 0.01 kg
n	≤ 3 000			
Min	10 g	20 g	0.04 kg	0.1 kg
Maximum Tara-balancing	3 000 g	6 000 g	15 kg	30 kg
Temperature range	-10 °C / +40 °C			
Power supply	Dry battery			

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Evaluation Report is not permitted, although either may be reproduced in full.