

# Agency for Technology & Standard, MOCIE

Member State of OIML Republic of Korea

OIML Certificate N° R76/1992-KR1-05.01

### OIML CERTIFICATE OF CONFORMITY

#### **Issuing authority**

: Agency for Techology & Standards, MOCIE

Address

: #2, Jungang-dong, Gwacheon-si, Gyeonggi-do, 427-716,

Republic of Korea

Person responsible : Na, Gi Hyoung - Researcher(Type approvals)

#### **Applicant**

Name

: BIOSPACE Co., Ltd

: #823, Yeoksam 1-domg, Gangnam-gu, 135-784

Seoul, Republic of Korea

#### Manufacturer of the certificated pattern

Name

: BIOSPACE Co., Ltd

Address

: #823, Yeoksam 1-domg, Gangnam-gu, 135-784

Seoul, Republic of Korea

## Identification of the certificated pattern

Type: INBODY 720

 $Max \leq 150 \text{ kg}$ 

e≥100 g

n≤1500 divisions

Temperature range :  $0^{\circ}$ C/+ $40^{\circ}$ C

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 76 edition 1992, including amendment 1(1994) for accuracy class III



# Agency for Technology & Standard, MOCIE

Member State of OIML Republic of Korea OIML Certificate N° R76/1992-KR1-05.01

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report: R76/1992-KR1-05.01, that includes 50 pages.

The issuing authority

MA. Gu-Hyoung

Date: 7. October. 2005

The CIML member

Jun, Yoo Tae

Date: 7. October. 2005

#### Important note:

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.