



Korean Agency for Technology & Standards, MOCIE

Member State of OIML  
Republic of Korea



OIML Certificate N°  
R76/1992-KR1-07.01

## OIML MAA CERTIFICATE OF CONFORMITY

### Issuing authority

Name : Korean Agency for Technology & Standards, MOCIE  
Address : #2, Jungang-dong, Gwacheon-si, Gyeonggi-do, 427-716, Republic of Korea  
Person responsible : Kim, Hyun Tae - Researcher(Type approvals)

### Applicant

Name : BIOSPACE Co., Ltd  
Address : #518-10, Dogok 2-dong, Gangnam-gu, 135-082 Seoul, Republic of Korea

### Manufacturer of the certificated pattern

Name : BIOSPACE Co., Ltd  
Address : #518-10, Dogok 2-dong, Gangnam-gu, 135-082 Seoul, Republic of Korea

### Identification of the certificated pattern

Type : INBODY 520  
Class : Accuracy Class Ⅲ  
Max :  $\text{Max} \leq 150 \text{ kg} (\text{e} \geq 100 \text{ g}, n \leq 1500 \text{ divisions}, n \leq 1500 \text{ divisions},$   
Temperature range :  $0^\circ\text{C}/+40^\circ\text{C}$ )

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R 76 edition 1992, including amendment 1(1994)  
for accuracy class Ⅲ

**Member State of OIML**  
**Republic of Korea**

**OIML Certificate N°**  
**R76/1992-KR1-07.01**

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation.

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report : R76/1992-KR-06.01 that includes 56 pages.

**The Issuing Authority**

**The OIML Member**

Mr. Kim, Hyun Tae  
15 January 2007

Dr. Jun, Yoo Tae  
15 January 2007

**Important note :**

Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.