

### Nederlands Meetinstituut

## OIML Certificate N° R76/1992-NL1-03.38 revision 1

# **OIML Member State**The Netherlands

Project number 503293 Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

**Issuing Authority** 

Name:

NMi Certin B.V.

Address:

Hugo de Grootplein 1, Dordrecht

Person responsible:

Ing. C. Oosterman

**Applicant** 

Name:

Ishida Co., Ltd.

Address:

44 Sanno-cho, Shogoin

Sakyo-ku, Kyoto -city, 606-8392

Japan

Manufacturer of the certified type

Name:

Ishida Co., Ltd.

Address:

44 Sanno-cho, Shogoin

Sakyo-ku, Kyoto -city, 606-8392

Japan

Identification of the certified type

Type: IPC, IPC-WP, JC or JC-WP

 $1.5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ 

e ≥ 1 g

 $n \le 3000$  divisions for single interval instruments

 $n \leq 1500 \ divisions$  (per partial weighing range) for multi interval instruments

maximum of two partial weighing ranges

 $T \le - Max_1$ 

Temperature range -5 °C / +40 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6076 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76** 

Edition 1992

for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

Nederlands Meetinstituut Hugo de Grootplein 1 3314 EG Dordrecht Telephone +31 78 6332332

Telefax +31 78 6332309

NMi B.V.

(Chamber of Commerce no.27.228.701)

Subsidiary companies:

NMi Van Swinden Laboratorium B.V. (27228703) NMi Certin B.V. (27.233.418) Verispect B.V. (27.228.700) This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



#### Nederlands Meetinstituut

## OIML Certificate N° R76/1992-NL1-03.38 revision 1

**OIML Member State**The Netherlands

Project number 503293 Page 2 of 2

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R76/1992-NL-02.23, that includes 55 pages;
- N° R76/1992-NL1-03.38 that includes 21 pages.

This revision replaces the earlier version.

The Issuing Authority

Ing. C. Obsterman Manager Product Certification

√22 April 2005

The CIML Member Dr. Ir C.J. van Mullem

22 April 2005

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.