

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Mettler-Toledo (Albstadt) GmbH  
Address: Unter dem Malesfelsen 34  
D-72458 Albstadt  
Germany

### Manufacturer of the certified type

Name: Mettler-Toledo (Albstadt) GmbH  
Address: Unter dem Malesfelsen 34  
D-72458 Albstadt  
Germany

### Identification of certified type

Non-automatic weighing instrument  
Type: Viper... and BB4 4...  
 $1.5 \text{ kg} \leq \text{Max} \leq 60 \text{ kg}$   
For the additional load receptor  $1.5 \text{ kg} \leq \text{Max} \leq 6000 \text{ kg}$   
 $e \geq 0.2 \text{ g}$   
 $n \leq 7500 \text{ divisions}$   
 $n \leq 5000 \text{ divisions (per weighing range)}$   
Maximum of 2 partial weighing ranges  
Temperature range  $-10^\circ\text{C} / +40^\circ\text{C}$



This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T5508 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class **II**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R76/1992-NL-99.07, that includes 52 pages;

N° R76/1992-NL-02.06, that includes 16 pages;

N° R76/1992-NL1-04.33A, that includes 10 pages;

N° R76/1992-NL1-04.33B, that includes 10 pages.

The Issuing Authority NL1  
NMI Certin, 24 February 2010



C. Oosterman  
Head Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.