

Nederlands Meetinstituut

OIML Certificate N° R76/1992-NL1-05.05

Project number 501509

Page 1 of 2

OIML Member State
The Netherlands

OIML CERTIFICATE OF CONFORMITY

Issuing Authority

Name: NMi Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: Ing. C. Oosterman

Applicant

Name: Avery Weigh-Tronix
Address: Foundry Lane, Smethwick
West Midlands, B66 2LP
United Kingdom

Manufacturer of the certified type

Name: Avery Weigh-Tronix
Address: Foundry Lane, Smethwick
West Midlands, B66 2LP
United Kingdom

Identification of the certified type

Type : HL122, L122, H205, H303, H305 & H400

$3 \text{ kg} \leq \text{Max} \leq 3000 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 3500 \text{ divisions}$

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6082 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

This Certificate does not bestow any form of legal international approval.


Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht
Telephone +31 78 6332332
Telefax +31 78 6332309

NMi B.V.
(Chamber of Commerce no.27.228.701)

Subsidiary companies:
NMi Van Swinden Laboratorium B.V. (27228703)
NMi Certin B.V. (27.233.418)
Verispect B.V. (27.228.700)

This document is issued under the provision that NMi. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

OIML Member State
The Netherlands

OIML Certificate N° R76/1992-NL1-05.05
Project number 501509
Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 201367, that includes 42 pages;

N° 205863, that includes 14 pages;

N° AB01711 (R60/1991-GB-96.01), that includes 24 pages;

N° AB01931 (R60/1991-GB-97.02), that includes 23 pages;

N° AB02070 (R60/1991-GB-99.04), that includes 20 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

14 February 2005

The OIML Member
Dr. ir. C.J. van Mullem

14 February 2005

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.