

OIML Member State  
The Netherlands

### OIML CERTIFICATE OF CONFORMITY

#### Issuing Authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1, Dordrecht  
Person responsible: Ing. C. Oosterman

#### Applicant

Name: Shinko Denshi Co., Ltd.  
Address: 3-9-11 Yushima, Bunkyo-ku  
Tokyo 113-0034  
Japan

#### Manufacturer of the certified type

Name: Shinko Denshi Co., Ltd.  
Address: 3-9-11 Yushima, Bunkyo-ku  
Tokyo 113-0034  
Japan

#### Identification of the certified type

Type : AJ(H), AJ for carat (CT), SJ or SJP

Class <b>I</b>	Class <b>II</b>
$n \leq 82000$ divisions	$n \leq 42000$ divisions
$500 \text{ g} \leq \text{Max} \leq 6200 \text{ g}$ or $2500 \text{ ct} \leq \text{Max} \leq 31000 \text{ ct}$	$50 \text{ g} \leq \text{Max} \leq 12 \text{ kg}$ or $250 \text{ ct} \leq \text{Max} \leq 60000 \text{ ct}$
$e \geq 0.01 \text{ g}$ or $e \geq 0.1 \text{ ct}$	$e \geq 0.01 \text{ g}$ $e \geq 0.1 \text{ ct}$
$e = d$ or $e = 10d$	$e = d$ or $e = 10d$

Temperature range +10 °C / +30 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6124 and the appertaining documentation folder), with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992

for accuracy class **I** and **II**


Nederlands Meetinstituut  
Hugo de Grootplein 1  
3314 EG Dordrecht  
Telephone +31 78 6332332  
Telefax +31 78 6332309

NMI B.V.  
(Chamber of Commerce no.27.228.701)

Subsidiary companies:  
NMI Van Swinden Laboratorium B.V. (27228703)  
NMI Certin B.V. (27.233.418)  
Verispect B.V. (27.228.700)

This document is issued under the provision that NMI. B.V. nor its subsidiary companies accept any liability.

Reproduction of the complete document is allowed. Parts of the document may only be reproduced after written permission



Nederlands Meetinstituut

**OIML Certificate N° R76/1992-NL1-05.09**

Project number 501358

Page 2 of 2

**OIML Member State**

The Netherlands

This Certificate relates only to the metrological and technical characteristics of the type of the instrument covered by the relevant OIML International Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated test reports:

N° R76/1992-NL1-05.09 that includes 15 pages;  
N° R76/1992-NL1-04.02 that includes 11 pages;  
N° R76/1992-NL1-03.45A that includes 47 pages;  
N° R76/1992-NL1-03.45B that includes 23 pages;  
N° R76/1992-NL-02.40A, that includes 38 pages;  
N° R76/1992-NL-02.40B, that includes 39 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

b/a 

31 March 2005

The OIML Member  
Dr. Ir. C.J. van Mullem

  
31 March 2005

\*  
\* \*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.