



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Shanghai Teraoka Electronic Co., LTD
Address: Tinglin Industry Developmental Zone
Jin Shan Country
Shanghai 201505
P.R. of China

Manufacturer of the certified type


Name: Shanghai Teraoka Electronic Co., LTD
Address: Tinglin Industry Developmental Zone
Jin Shan Country
Shanghai 201505
P.R. of China

Identification of certified type

Non-automatic weighing instrument
Family of type: DS-650...

Instrument with load cell type K	Instrument with load cell type P
$6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ or $\text{Max} \leq 30 \text{ lb}$	$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ or $\text{Max} \leq 60 \text{ lb}$
$e \geq 2 \text{ g}$	$e \geq 1 \text{ g}$

$n \leq 3000$ divisions (per partial weighing range)
Maximum of two partial weighing ranges
 $T \leq - \text{Max}_1$
Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-06.03

Project number 507860

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T2214 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL-93.11 that includes 58 pages;

N° R76/1992-NL-94.04A that includes 58 pages;

N° R76/1992-NL-94.04B that includes 58 pages;

N° R76/1992-NL1-06.03 that includes 54 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

24 January 2006

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.