



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Ohaus Corporation
Address: 19A Chapin Road
Pine Brook, NJ 07058-9878
United States of America


Manufacturer of the certified type

Name: Ohaus Corporation
Address: 19A Chapin Road
Pine Brook, NJ 07058-9878
United States of America

Identification of certified type

Non-automatic weighing instrument
Type: Voyager and Explorer (Pro)

Class **I**
 $62 \text{ g} \leq \text{Max} \leq 210 \text{ g}$
 $e \geq 1 \text{ mg}$
 $e = d \text{ or } e = 10 d$
 $n \leq 210000 \text{ divisions}$



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-06.11

Project number 503855

Page 2 of 2

Class **II**
 $210 \text{ g} \leq \text{Max} \leq 8100 \text{ g}$
 $e \geq 10 \text{ mg}$
 $e = d \text{ or } e = 10 \text{ d}$
 $n \leq 81000 \text{ divisions}$

Class **II**
 $12 \text{ kg} \leq \text{Max} \leq 32.1 \text{ kg}$
 $e \geq 1 \text{ g}$
 $e = d \text{ or } e = 10 \text{ d}$
 $n \leq 32100 \text{ divisions}$

Temperature range +10 °C / +30 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T2914 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class **I** and **II**


This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- N° R76/1992-NL-97.09A test report that includes 48 pages;
- N° R76/1992-NL-97.09B test report that includes 48 pages;
- N° R76/1992-NL1-06.11 that includes 39 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


2 March 2006

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.