



**OIML Member state**  
The Netherlands

**OIML Certificate N° R76/1992-NL1-07.03**  
Project number 608445  
Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Shanghai Teraoka Electronic Co.,LTD.  
Address: Tinglin Industry Development Zone  
201505 Shanghai  
PR of China

### Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co.,LTD.  
Address: Tinglin Industry Development Zone  
201505 Shanghai  
PR of China

### Identification of certified type

Non-automatic weighing instrument  
Type : SM-100..

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$   
 $e \geq 1 \text{ g}$   
 $n \leq 3000$  divisions (per partial weighing range)  
maximum of 2 partial weighing ranges  
 $T \leq - \text{Max}_1$  for multi-interval scales

Temperature range -10 °C / +40 °C



**OIML Member state**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6734 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-05.11 that includes 57 pages;

N° R76/1992-NL1-05.18 that includes 15 pages;

N° R76/1992-NL1-04.13A that includes 56 pages;

N° R76/1992-NL1-05.31A that includes 24 pages;

N° R76/1992-NL1-05.31B that includes 16 pages.

N° R76/1992-NL1-05.40 that includes 16 pages.

N° R76/1992-NL1-07.03 that includes 43 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

2 February 2007

\*

\*\*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.