



## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Shanghai Teraoka Electronic Co., Ltd.  
Address: Shanghai Teraoka Electronic Co., Ltd.  
Tinglin industry developmental zone  
Jin Shan county Shanghai 201505  
P.R. of China

### Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co., Ltd.  
Address: Shanghai Teraoka Electronic Co., Ltd.  
Tinglin industry developmental zone  
Jin Shan county Shanghai 201505  
P.R. of China

### Identification of certified type

Non-automatic weighing instrument  
Type: DS-162, DS-162SS, DS-162C  
  
 $6 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$   
 $e \geq 1 \text{ g}$   
 $n \leq 7500$  divisions (per partial weighing range)  
maximum of 2 partial weighing ranges  
 $T \leq - \text{Max}_1$  for multi-interval  
 $T \leq - 50\%$  of Max for single-interval  
  
Temperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



**OIML Member state**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7013 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class  $\textcircled{\text{III}}$  or  $\textcircled{\text{III}}$

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 600790 that includes 48 pages;

R60/1991-NL-97.14, That includes 37pages (Load cell type P);

R60/1991-NL-99.09, that includes 36pages (Load cell type P);

R60/1991-NL-97.23A, that includes 37pages (Load cell type N);

R60/1991-NL-97.23B, that includes 37pages (load cell type N);

R60/1991-NL-97.18A, that includes 37pages (load cell type K);

R60/1991-NL-97.18B, that includes 37pages (load cell type K);

R60/1991-NL-95.07, that includes 38 pages (Load cell type M).

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

13 February 2007

\*

\*\*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.