



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.
Address: No. 10 Kunlan Road,
ChangZhou XinBei District
213125 Jiangsu
Peoples Republic of China

Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Measurement Technology Ltd.
Address: No. 10 Kunlan Road,
ChangZhou XinBei District
213125 Jiangsu
Peoples Republic of China

Identification of certified type

Non-automatic weighing instrument
Type: RN10...(Tiger II)

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ or $6 \text{ lb} \leq \text{Max} \leq 60 \text{ lb}$
 $e \geq 1 \text{ g}$
 $n \leq 3000$ divisions

Temperature range: -10°C / $+40^{\circ}\text{C}$



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6234 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R76/1992-NL1-03.11, that includes 56 pages;

N° R76/1992-NL1-07.13 that includes 36 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification


24 May 2007

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.