



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Ohaus Corporation
Address: 19A Chapin Road
Pine Brook
New Jersey 07058
United States of America

Manufacturer of the certified type

Name: Ohaus Corporation
Address: 19A Chapin Road
Pine Brook
New Jersey 07058
United States of America

Identification of certified type

Non-automatic weighing instrument
Type: Valor 3000 series

For load cells of type 72198183:
Max = 300 g
 $e \geq 0.2$ g
 $n \leq 1500$ divisions

For load cells of type 72193366:
 $3 \text{ kg} \leq \text{Max} \leq 6 \text{ kg}$
 $e \geq 1$ g
 $n \leq 3000$ divisions

$T \leq - \text{Max}$

Temperature range 0 °C / +40 °C



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7125 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-07.18A that includes 52 pages.

N° R76/1992-NL1-07.18B that includes 34 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

1/a

5 June 2007

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.