



## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: OHAUS Corporation  
Address: 19A Chapin Road  
Pine Brook, NJ 07058  
USA

### Manufacturer of the certified type

Name: OHAUS Corporation  
Address: 19A Chapin Road  
Pine Brook, NJ 07058  
USA

### Identification of certified type

Non-automatic weighing instrument  
Type: Trooper

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$   
 $e \geq 1 \text{ g}$   
 $n \leq 3000 \text{ divisions}$   
 $T \leq - \text{Max}$

Temperature range  $0 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



**OIML Member state**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6028 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class (III)


This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R76/1992-NL1-02.10 that includes 52 pages;
- N° R76/1992-NL1-07.13 that includes 36 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

  
14 June 2007

\*

\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.