



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Tanita Corporation
Address: 14-2, 1 Chome, Maeno-Cho
Itabashi-Ku, Tokyo
Japan

Manufacturer of the certified type

Name: Tanita Corporation
Address: 14-2, 1 Chome, Maeno-Cho
Itabashi-Ku, Tokyo
Japan

Identification of certified type

Non-automatic weighing instrument
Type: BD-815MA

$6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$
 $e \geq 2 \text{ g}$
 $n \leq 3000$ divisions (per partial weighing range)
maximum of 2 partial weighing ranges
 $T \leq - (\text{Max}-e)$

Temperature range $+5^\circ\text{C} / +35^\circ\text{C}$



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-03.37

Project number 702404

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7304 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-07.37A that includes 51 pages;

N° R76/1992-NL1-07.37B that includes 33 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

2 November 2007

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.