

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Person responsible: Ing. C. Oosterman

### Applicant

Name: Shanghai Teraoka Electronic Co., Ltd.

Address: Tinglin Industry Developmental Zone  
Jin Shan District  
Shanghai 201505  
Peoples Republic of China

### Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co., Ltd.

Address: Tinglin Industry Developmental Zone  
Jin Shan District  
Shanghai 201505  
Peoples Republic of China

### Identification of certified type

Non-automatic weighing instrument  
Type: DS-673(SS) $1.5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$  $e \geq 0.5 \text{ g}$  $n \leq 7500$  divisions for single interval instruments $n \leq 3000$  divisions for multi interval instruments (per partial weighing range)

Maximum of 2 partial weighing ranges

 $T \leq -50\%$  for single interval instruments $T \leq -\text{Max}_1$  for multi interval instrumentsTemperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$  for DS-673 and  $0^\circ\text{C} / +40 \text{ }^\circ\text{C}$  for DS-673SS



This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7313 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
Edition 1992 (E)  
for accuracy class (III)


This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-07.38 that includes 56 pages;  
N° R76/1992-NL1-07.38b that includes 28 pages.

The Issuing Authority, NL1  
NMI Certin, 19 November 2009

  
C. Oosterman  
Head Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.