



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
 Address: Hugo de Grootplein 1
 3314 EG Dordrecht
 The Netherlands
 Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo (Albstadt) GmbH
 Address: Unter dem Malesfelsen 34
 D-72458 Albstadt
 Germany

Manufacturer of the certified type

Name: Mettler-Toledo (Albstadt) GmbH
 Address: Unter dem Malesfelsen 34
 D-72458 Albstadt
 Germany

Identification of certified type

Non-automatic weighing instrument
 Type: BBK4...and VIPER...

II	III multi-interval
$n \leq 61000$ divisions	$n \leq 10000$ divisions (per weighing range)
$610 \text{ g} \leq \text{Max} \leq 35.1 \text{ kg}$	$1.5 \text{ kg} \leq \text{Max} \leq 60 \text{ kg}$ For the additional load receptor $1.5 \text{ kg} \leq \text{Max} \leq 6000 \text{ kg}$
$e \geq 10 \text{ mg}$	$e \geq 1 \text{ g}$
$e = d$ or $e = 10d$	$e = d$
$+10 \text{ }^\circ\text{C} / +30 \text{ }^\circ\text{C}$	$+5 \text{ }^\circ\text{C} / +35 \text{ }^\circ\text{C}$

NMI Certin BV
 Hugo de Grootplein 1, 3314 EG Dordrecht
 PO Box 394, 3300 AJ Dordrecht, NL
 T +31 78 6332332
 F +31 78 6332309
 certin@nmi.nl
 www.nmi.nl

This document is issued under the provision
 that no liability is accepted and that the
 applicant shall indemnify third-party liability.

Reproduction of the complete document is
 permitted.
 The notification of NMI Certin as Issuing
 Authority can be verified at www.oiml.org.





This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T5627 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class (II) or (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

- No R76/1992-NL-00.22A, that includes 52 pages;
- No R76/1992-NL-00.22B, that includes 31 pages;
- No R76/1992-NL-01.23A, that includes 39 pages;
- No R76/1992-NL-01.23B, that includes 10 pages;
- No R76/1992-NL1-04.32, that includes 30 pages;
- No R76/1992-NL1-08-12, that includes 31 pages.

The Issuing Authority NL1
NMI Certin, 24 February 2010


C. Oosterman
Head Certification Board

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.