



**OIML Member state**  
The Netherlands

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Mettler-Toledo (Changzhou) Scale & System Ltd.  
Address: No. 10 Kunlan Road,  
ChangZhou XinBei District  
213125 Jiangsu  
Peoples Republic of China

### Manufacturer of the certified type

Name: Mettler-Toledo (Changzhou) Scale & System Ltd.  
Address: No. 10 Kunlan Road,  
ChangZhou XinBei District  
213125 Jiangsu  
Peoples Republic of China



OIML Member state  
The Netherlands

Identification of certified type

Non-automatic weighing instrument  
Type: 8442 (Tiger P)

AMI load cell:  
15 kg ≤ Max ≤ 30 kg  
e ≥ 2 g  
n ≤ 3000 divisions (per  
partial weighing range)

AMI (New) load cell:  
6 kg ≤ Max ≤ 30 kg  
e ≥ 1 g  
single interval: n ≤ 6000 divisions  
multi-interval: n ≤ 3000 divisions  
(per partial weighing range)

maximum of two partial weighing ranges  
T ≤ - (Max<sub>1</sub>-e)  
Temperature range -10 °C / +40 °C

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6311 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
Edition 1992 (E)  
for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

- N° R76/1992-NL1-06.08 that includes 56 pages;
- N° R76/1992-NL1-06.18A that includes 44 pages;
- N° R76/1992-NL1-06.18B that includes 28 pages;
- N° R76/1992-NL1-08.21 that includes 12 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

25 June 2008

\*  
\* \*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.