

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.

Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands

Person responsible: Ing. C. Oosterman

### Applicant

Name: Taiwan Scale Mfg. Co., Ltd.

Address: 282, Sec. 3  
Hoping W. Road  
Taipei  
Taiwan

### Manufacturer of the certified type

Name: Taiwan Scale Mfg. Co., Ltd.

Address: 282, Sec. 3  
Hoping W. Road  
Taipei  
Taiwan

### Identification of certified type

Non-automatic weighing instrument  
Type: ASP / ATP / AHW / AHC / QSP / QTP / QHW / QHC

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 3000$  divisions (per partial weighing range)

Maximum of 2 partial weighing ranges

$T \leq -\text{Max}$

Temperature range  $-10^\circ\text{C} / +40^\circ\text{C}$

#### NMI Certin BV

Hugo de Grootplein 1, 3314 EG Dordrecht  
PO Box 394, 3300 AJ Dordrecht, NL  
T +31 78 6332332  
F +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

Reproduction of the complete document is permitted.  
The notification of NMI Certin as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org).

**OIML Member state**

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7305 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class III

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° R76/1992-NL1-07.35A that includes 53 pages;

N° R76/1992-NL1-07.35B that includes 14 pages

N° R76/1992-NL1-08.03A that includes 14 pages;

N° R76/1992-NL1-08.03B that includes 12 pages;

N° R76/1992-NL1-08.22 that includes 12 pages;

N° R76/1992-NL1-08.22B that includes 15 pages.

The Issuing Authority NL1  
NMI Certin, 30 October 2009



C. Oosterman  
Head Certification Board

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.