



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-08.29

Project number 805851

Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Excell Precision Co. Ltd.
Address: 6F, No. 127, Lane 235
Pao-Chiao Road, Hsin Tien
Taipei Hsien, Taiwan, R.O.C.

Manufacturer of the certified type

Name: Excell Precision Co. Ltd.
Address: 6F, No. 127, Lane 235
Pao-Chiao Road, Hsin Tien
Taipei Hsien, Taiwan, R.O.C.

Identification of certified type

Non-automatic weighing instrument
Type: AW, GW, GWH, FA130, FD130, FG130 & SA130

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$
 $e \geq 1 \text{ g}$
 $n \leq 3500$ divisions (per partial weighing range)
maximum of two partial weighing ranges

Temperature range -10 °C / +40 °C



OIML Member state

The Netherlands

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6211 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 201367, that includes 42 pages;

N° 215140, that includes 24 pages;

N° 215140a, that includes 15 pages;

N° 601459, that includes 16 pages;

N° R60/1991-NL-96.03, that includes 37 pages;

N° R60/1991-NL-96.05, that includes 35 pages;

N° R60/2000-NL-02.21B, that includes 37 pages;

N° R60/1991-NL-97.30, that includes 40 pages;

N° R60/2000-NL-00.09A, that includes 40 pages;

N° R60/2000-NL-00.09B, that includes 12 pages.

The Issuing Authority

Ing. C. Oosterman

Manager Product Certification

14 July 2008

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.