



**OIML Member state**  
The Netherlands

**OIML Certificate N° R76/1992-NL1-08.34**

Project number 803891

Page 1 of 2

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: TERAOKA WEIGH-SYSTEM PTE LTD.  
Address: 4 Leng Kee Road  
#06-01 SIS Building  
SINGAPORE 159088

### Manufacturer of the certified type

Name: TERAOKA WEIGH-SYSTEM PTE LTD.  
Address: 4 Leng Kee Road  
#06-01 SIS Building  
SINGAPORE 159088

### Identification of certified type

Non-automatic weighing instrument  
Type: SM-800.. / SM-880..

$6 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$   
 $30 \text{ kg} \leq \text{Max} \leq 300 \text{ kg}$  (instruments with external weighing platform)  
 $e \geq 1 \text{ g}$   
 $e \geq 5 \text{ g}$  (instruments with external weighing platform)  
 $n \leq 3000$  divisions (per partial weighing range)  
Maximum of two partial weighing ranges  
 $T \leq - \text{Max}_1$

Temperature range  $-10 \text{ °C} / +40 \text{ °C}$

**NMI Certin B.V.**  
Hugo de Grootplein 1, 3314 EG Dordrecht  
P.O. Box 394, 3300 AJ Dordrecht, NL  
phone +31 78 6332332  
fax +31 78 6332309  
certin@nmi.nl  
www.nmi.nl

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMI B.V. (see "Regulation objection and appeal against decisions of NMI B.V.")

NMI Certin B.V., chamber o.c. nr. 27.233.418

This document is issued under the provision that no responsibility is accepted and that the applicant gives warranty for each responsibility against third parties.

The notification of NMI Certin as Issuing Authority can be verified at [www.oiml.org](http://www.oiml.org).



**OIML Member state**  
The Netherlands

**OIML Certificate N° R76/1992-NL1-08.34**

Project number 803891  
Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T6371 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**  
Edition 1992 (E)  
for accuracy class **III**

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-04.01A that includes 56 pages;  
N° R76/1992-NL1-04.01B that includes 12 pages;  
N° R76/1992-NL1-04.26 that includes 16 pages;  
N° R76/1992-NL1-06.34A that includes 42 pages;  
N° R76/1992-NL1-06.34B that includes 35 pages;  
N° R76/1992-NL1-08.34 that includes 24 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

3 September 2008

\*  
\* \*

**Important note:** Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.