



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin Industry Developmental Zone
Jin Shan District
Shanghai 201505
Peoples Republic of China

Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin Industry Developmental Zone
Jin Shan District
Shanghai 201505
Peoples Republic of China

Identification of certified type

Non-automatic weighing instrument
Type: DS-671(II)(H)

$1.5 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 0.5 \text{ g}$

$n \leq 3000 \text{ divisions}$

$T \leq -\text{Max}$

Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



OIML Member state
The Netherlands

Project number 806167
Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T5720 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76
Edition 1992 (E)
for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL-01.20, that includes 14 pages;
N° R76/1992-NL-01.21, that includes 32 pages;
N° R76/1992-NL-00.11, that includes 52 pages;
N° R76/1992-NL1-08.39A that includes 42 pages;
N° R76/1992-NL1-08.39B that includes 12 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification


20 October 2008

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.