



**OIML Member state**  
The Netherlands

## OIML CERTIFICATE OF CONFORMITY

### Issuing authority

Name: NMI Certin B.V.  
Address: Hugo de Grootplein 1  
3314 EG Dordrecht  
The Netherlands  
Person responsible: Ing. C. Oosterman

### Applicant

Name: Adam Equipment Co., Ltd.  
Address: Bond Avenue, Denbigh East Industrial Estate  
Milton Keynes, MK1 1SW  
United Kingdom

### Manufacturer of the certified type

Name: Adam Equipment Co., Ltd.  
Address: Bond Avenue, Denbigh East Industrial Estate  
Milton Keynes, MK1 1SW  
United Kingdom

### Identification of certified type

Non-automatic weighing instrument  
Type: WBW ..M / WBZ ..

$1.5 \text{ kg} \leq \text{Max} \leq 15 \text{ kg}$   
 $e \geq 0.5 \text{ g}$   
 $n \leq 3000 \text{ divisions}$

$T \leq - \text{Max}$  for WBW and  $T < -10 \text{ kg}$  for WBZ instruments with  $\text{Max} \geq 10 \text{ kg}$

Temperature range  $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7476 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

**R76**

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° 708724A that includes 46 pages;

N° 708724B that includes 25 pages;

N° R60/1991-NL-97.30, that includes 40 pages;

N° R60/2000-NL-00.09A, that includes 40 pages;

N° R60/2000-NL-00.09B, that includes 12 pages;

N° 2000-WJ-0016 that includes 18 pages;

N° 2000-WJ-0018 that includes 18 pages.

The Issuing Authority  
Ing. C. Oosterman  
Manager Product Certification

5 November 2008

\*

\*\*

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.