



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin Industry Developmental Zone
Jin Shan District
Shanghai 201505
Peoples Republic of China

Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin Industry Developmental Zone
Jin Shan District
Shanghai 201505
Peoples Republic of China

Identification of certified type

Non-automatic weighing instrument
Type: DS-781, DS-782

$3 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$

$e \geq 1 \text{ g}$

$n \leq 6000$ divisions for single-interval

$n \leq 3000$ divisions (per partial weighing range) for multi-interval
maximum of 2 partial weighing ranges

$T \leq -\frac{1}{2} \text{ Max}$ for single-interval instruments

$T \leq -\text{Max}_1 - e_1$ for multi-interval instruments

Temperature range $-10 \text{ }^\circ\text{C} / +40 \text{ }^\circ\text{C}$



Nederlands Meetinstituut

OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-08.43

Project number 809821

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7475 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-08.38A that includes 57 pages;

N° R76/1992-NL1-08.38B that includes 43 pages;

N° R76/1992-NL1-08.43 that includes 43 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

22 December 2008

*

* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.