



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin industry developmental zone
Jin Shan county Shanghai 201505
P.R. of China

Manufacturer of the certified type

Name: Shanghai Teraoka Electronic Co., Ltd.
Address: Tinglin industry developmental zone
Jin Shan county Shanghai 201505
P.R. of China

Identification of certified type

Non-automatic weighing instrument
Type: DS-620

$6 \text{ kg} \leq \text{Max} \leq 600 \text{ kg}$
 $e \geq 1 \text{ g}$
 $n \leq 7500$ divisions (per partial weighing range)
maximum of 2 partial weighing ranges
 $T \leq - \text{Max}_1$ for multi-interval
 $T \leq - 50\%$ of Max for single-interval

Temperature range $-10 \text{ °C} / +40 \text{ °C}$



This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7508 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III) or (IIII)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Reports:

N° 807678 that includes 44 pages;

R60/1991-NL-97.14, that includes 37 pages (load cell type P);

R60/1991-NL-99.09, that includes 36 pages (load cell type P);

R60/1991-NL-97.23A, that includes 37 pages (load cell type N);

R60/1991-NL-97.23B, that includes 37 pages (load cell type N);

R60/1991-NL-97.18A, that includes 37 pages (load cell type K);

R60/1991-NL-97.18B, that includes 37 pages (load cell type K);

R60/1991-NL-95.07, that includes 38 pages (load cell type M).

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

4 December 2008

*

**

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.