



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Mettler-Toledo AG
Address: Im Langacher
8606 Greifensee
Switzerland

Manufacturer of the certified type

Name: Mettler-Toledo AG
Address: Im Langacher
8606 Greifensee
Switzerland

Identification of certified type

Non-automatic weighing instrument
Type: WX series

$110 \text{ g} \leq \text{Max} \leq 220 \text{ g}$ or the equivalent in ct
 $e \geq 1 \text{ mg}$ or the equivalent in ct
 $d \leq e \leq 100 d$
 $n \leq 220000$ divisions
 $T \leq - \text{Max}$

Temperature range +10 °C / +30 °C



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-08.45

Project number 808920

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7494 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class $\textcircled{\text{I}}$

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:
N° R76/1992-NL1-08.45 that includes 39 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

1/0 

23 December 2008

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.