



OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands
Person responsible: Ing. C. Oosterman

Applicant

Name: Fook Tin Technologies Ltd.
Address: Eastern Centre, 1065 King's Rd
Quarry Bay
Hong Kong

Manufacturer of the certified type

Name: Fook Tin Technologies Ltd.
Address: Eastern Centre, 1065 King's Rd
Quarry Bay
Hong Kong

Identification of certified type

Non-automatic weighing instrument
Type: FMM-T370x / TPRO2x00

Max ≤ 15 kg
e ≥ 5 g
n ≤ 1500 divisions (per partial weighing range)
Maximum of 2 partial weighing ranges

Temperature range 10°C / 40°C



OIML Member state
The Netherlands

OIML Certificate N° R76/1992-NL1-09.05

Project number 808263

Page 2 of 2

This Certificate attests the conformity of the above identified type (represented by the sample or samples identified in the associated Test Report, the type-approval certificate and the description with number T7537 and the appertaining documentation folder) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R76

Edition 1992 (E)

for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of instrument covered by the relevant OIML Recommendation identified above.

This Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated Test Report:

N° R76/1992-NL1-09.05 that includes 52 pages.

The Issuing Authority
Ing. C. Oosterman
Manager Product Certification

20 February 2009

*
* *

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated Test Report is not permitted, although either may be reproduced in full.